

AMENDMENT NO.3 Dated 15.10.2026

REF TENDER: HLL/CHO/PSD/HCS/2026-27/TENDER 06 Dated 19.09.2026

Amendment made to the details given in Tender Document.

REF NO: HLL/CHO/PSD/HCS/2026-27/TENDER 06 Dated 19.09.2026

CPP TENDER ID : 2026_HLL_291473_1

The following amendment has been incorporated in the above mentioned **E-Tender for the supply of Fully Automatic Medium Throughput Immunoassay Analyser on Reagent Rental basis to HINDLABS Bhopal and for meeting future requirements at HINDLABS for a period of five years.**

The following amendment / Inclusion have been incorporated to the bid document for the above tender:

1. Notice Inviting Tender

FOR

LAST DATE AND TIME FOR RECEIPT OF TENDER – 17.10.2026, 15:30 Hrs

MAY BE READ AS

LAST DATE AND TIME FOR RECEIPT OF TENDER - 23.10.2026, 15:30 Hrs

2. Notice Inviting Tender

FOR

TIME AND DATE OF OPENING OF THE TENDER - 19.10.2026, 15:30 Hrs

MAY BE READ AS

TIME AND DATE OF OPENING OF THE TENDER – 24.10.2026, 15:30 Hrs

3. ANNEXURE 4 – TECHNICAL SPECIFICATION AND PARAMETERS

FOR

Point No: 6 - The analyzer should have a through-put of a minimum of 150 tests/hour.

MAY BE READ AS

Point No: 6 - The analyzer should have a through-put of 120 tests/hour or higher.

AMENDMENT NO.3 Dated 15.10.2026

REF TENDER: HLL/CHO/PSD/HCS/2026-27/TENDER 06 Dated 19.09.2026

4. ANNEXURE 4 – TECHNICAL SPECIFICATION AND PARAMETERS

FOR

Point No: 26 - Parameters to be offered.

Point d - Cardia Panel: CK-MB, Myoglobin, Troponin-I, BNP

MAY BE READ AS

Point No: 26 - Parameters to be offered.

Point d - Cardiac Panel: CK-MB, Myoglobin, Troponin-I or Troponin-T, BNP or NT-Pro BNP

5. ANNEXURE 4 – TECHNICAL SPECIFICATION AND PARAMETERS

FOR

Point No: 16 - The system should have a facility for storage and retrieval of at least 10,000 patient's samples.

MAY BE READ AS

Point No: 16 - The system should have a facility for storage and retrieval of at least 10,000 patient's samples through the analyser/software/LIS infrastructure, wherever technically appropriate.

6. BOQ

FOR

Sl.No. 36 and Item No 36 - Troponin-I

MAY BE READ AS

Sl.No. 36 and Item No 36 - Troponin-I/Troponin-T

7. BOQ

FOR

Sl.No. 37 and Item No 37 – BNP

AMENDMENT NO.3 Dated 15.10.2026

REF TENDER: HLL/CHO/PSD/HCS/2026-27/TENDER 06 Dated 19.09.2026

MAY BE READ AS

Sl.No. 37 and Item No 37 – BNP/NT-Pro BNP

The same amendment shall be applicable in the BOQ as well.

All relevant clauses of the Tender document are to be read in accordance with the above change and documents to be submitted are to be in compliance of the above. All other specifications, terms and conditions of the original tender document shall remain unchanged.

Vice President (PS) & Group Head (HCS)