Amendment No: 1 Issued on dated 07/11/2014

Ref: Tender Enquiry No: HLL/PCD/CDSCO-02/14-15 dtd.14.10.2014 and Pre-Bid meeting held on 27/10/2014 at HLL Lifecare Ltd., B-14A, Sector-62, Noida - 201 307

The following amendments are issued with respect to the above tender Enquiry Document.

SECTION I

NOTICE INVITING TENDERS (NIT)

(2) Tender No.: HLL/PCD/CDSCO-02/14-15

Existing:-

Sl No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	15.10.2014 to 14.11.2014
ii.	Place of sale of Tender Enquiry Documents	HLL Lifecare Ltd, (A Government of India Enterprise), Procurement & Consultancy Services Division, B-14 A, Sector-62, Noida -201 307
iii.	Cost of the Tender Enquiry Document	Rs. 3000/-
Iv	Pre Tender Meeting Date & Time	27.10.2014 at 1100 hrs.
V	Pre Tender Meeting Venue	Same as 2 (ii)
vi.	Closing date & time for receipt of Tender	15.11.2014 at 1400 hrs
vii.	Time and date of opening of Techno – Commercial tenders	15.11.2014 at 1430 hrs
Viii	Venue of Opening of Techno Commercial Tender	Same as above

Read As:-

Sl No.	Description	Schedule
i.	Dates of sale of tender enquiry documents	15.10.2014 to 24.11.2014
ii.	Place of sale of Tender Enquiry Documents	HLL Lifecare Ltd, (A Government of India Enterprise), Procurement & Consultancy Services Division, B-14 A, Sector-62, Noida -201 307
iii.	Cost of the Tender Enquiry Document	Rs. 3000/-
Iv	Pre Tender Meeting Date & Time	27.10.2014 at 1100 hrs.
V	Pre Tender Meeting Venue	Same as 2 (ii)
vi.	Closing date & time for receipt of Tender	25.11.2014 at 1400 hrs
vii.	Time and date of opening of Techno – Commercial tenders	25.11.2014 at 1430 hrs
Viii	Venue of Opening of Techno Commercial Tender	Same as above

Section – VII Technical Specifications

Item No. 1 Fluorescence Spectrophotometer

1. Added Para in the existing Technical specification:

Sensitivity: Signal to noise (S/N) ratio should be \geq 800 for Raman line of water (excitation wavelength 350 nm and 5 nm band width).

Item No. 3 Total Organic Carbon Analyzer

1. Existing:-Para: DOCUMENTS AND TRAININGS:

On site Calibration with traceable reference material, to be done by the Supplier on installation and there after every six months during warranty and CMC period.

Read as: - DOCUMENTS AND TRAININGS:

On site Calibration with traceable reference material, to be done by the Supplier on installation and there after every six months (the validity/stability of calibration should be six months) during warranty and CMC period

Item No. 6 Liquid Particle Counter

1. Existing: Para 1: Syringe Size: 1, 5 & 10 ml

Read as:- Syringe Size: 1ml & 10 ml

Item No. 10

Water Purification System

1. Added Para:

RO system specifications for feeding to ultrapure water system:

- 1. Make up rate of RO: 5 to 7 LPH
- 2. Inorganic rejection upto 98%
- 3.Organic rejection upto 99%
- 4. Total Organic carbon < 50 ppb
- 5.Bacteria < 5 cfu/ml

Amendment No.1

Item No. 8 <u>Digital Polarimeter</u>

1. Existing Para: 2.Optical rotation specifications a) Measuring Range: ± 89.9°:
Read as: a) Measuring Range: ± 89.0°
All other contents of the tender enquiry including terms & conditions remain unaltered.
End