

Amendment No. 4**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry HITES/PCD/PRECLINICAL/RC-01/18-19 dated 14/09//2018**

The following Technical changes are being incorporated in the above referred Tender Enquiry Document.

**Section VII
Technical Specification****Schedule No. 06
Biosafety Cabinet (Rfx No. 3000003478)**

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 56 Para 4	Dimensions (Cabinet Size): 6 feet. The interior of the cabinet shall be of stainless steel or equivalent, outer body powder coated steel material and must be smooth to ensure no risk of cuts to the users.	Dimensions (Cabinet Size): 6 feet or work surface area approx 1200 mm x 600 mm x 600 mm . The interior of the cabinet shall be of stainless steel or equivalent, outer body powder coated steel material and must be smooth to ensure no risk of cuts to the users.

**Schedule No. 09
Centrifuge (Rfx No. 3000003481)**

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 58 Para 5.1	Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility.	Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility / IEC-61010-2-020 .

**Schedule No. 37
Incubator (Rfx No. 3000003509)**

Tender Page No. & Para	TENDER SPECIFICATION	Read As
Pg 75 Para 5.1	Shall meet IEC-60601-1-2 :2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility.	Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility / IEC-61010-1 .

**Schedule No. 34
Human Bones set disarticulated (Rfx No. 3000003506)**

1	Set of real human bones of right and left sides having the following:
a	Shoulder girdle and upper limb bones- clavicle, scapula, humerus, radius, ulna, articulated hand with carpal bone
b	Pelvic girdle and lower limb bones- hip bone, femur, tibia, fibula, articulated foot and patella
c	Complete set of ribs and vertebrae and sacrum

d	Skull with mandible and individual bones (frontal, parietal, occipital, sphenoid and temporal)
2	The disarticulated adult bone set should be ideal for teaching the basics of human anatomy
3	It should be bleached and glazed and not easily brittle
4	Should contain pediatric, neonate and old age bones of male & female – 5 sets each.
5	Should contain adult disarticulated bones (male and female) – 30 each.
6	Should submit license for sale of human bones.

All other contents of the tender enquiry including terms & conditions remain unaltered.

Note:

- i. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**