

Amendment No. 4

08.11.19

Sub: Amendment to the Bidding Document

Ref.: Notice Inviting Bid ref. HITES/PCD/NCI-AIIMS/42/19-20 dated 30.08.2019 read with its Amendment No. 1, 2 & 3 dated 11.10.19, 23.10.19 & 24.10.19 respectively

The following changes have been authorised and are being incorporated in the above referred Bidding Document.

SECTION – VII**TECHNICAL SPECIFICATION AND GENERAL POINTS****A. TECHNICAL SPECIFICATION:****Item no. 1 (Tender ID: 2019 HLL 31185 2)****Blood Bank Equipment – Package 2****Item No. 16 Haemostatic Analyzer**

Tender Page & Para	Tender Specification	Amended as
Page 64; Para 11	It should have low molecular weight heparin management	Deleted

Item No. 24 Water Bath

Tender Page & Para	Tender Specification	Amended as
Page 71; Para 2	Should be top loading, have Rectangular, vertical chamber with well insulated jacket, chamber Volume minimum 450 liters or more.	Should be top loading, have Rectangular, vertical chamber with well insulated jacket, chamber Volume minimum 45 liters or more.

Item no. 8 (Tender ID: 2019 HLL 31185 8)**Mobile Blood Donation Van**

Tender Page & Para	Tender Specification	Amended as
Amndt 2; Page-23 Item no. 8	The net interior dimensions of MBDV body (blood collection area excluding the driver cabin) shall be min. 2000 mm. in width, 2000 mm. in height and min. 9500 mm. in length. The firm must provide expandable Donation compartment by at least 500 mm or more (width wise), of length 6-8 ft by hydraulics or electric pressure jack system during parked position which can safely retracted back to the normal width of the MBDV while in motion.	The net interior dimensions of MBDV body (blood collection area excluding the driver cabin) shall be min. 2000 mm. in width, 2000 mm. in height and min. 9500 mm. in length.

Important Note:

Technical Compliance Sheet 'Version 1.1' for Tender ID- **2019_HLL_31185_2, 2019_HLL_31185_8** under this NIB are being uploaded on CPPP. The same need to be downloaded for submission. This will supersede the previous version i.e. 'Ver1.0'.

All other contents of the Bidding Document including terms & conditions remain unaltered.