AMENDMENT NO.9 Dated 18.04.2020

Ref No: HLL/SD/RBD/2019-20/TENDER/15 Dt: 05.03.2020

Title: Essential Technical features for Ventilator for COVID 19

The following amendment has been incorporated to the bid document for the above tender;

Essential Technical features for Ventilator for COVID 19 as per amendment No. 07 dated 31.03.2020 stands cancelled and the revised specification is given below. All other terms and conditions as per amendment No. 06 dated 28.03.2020 remains same.

The Essential Technical features for Ventilator for COVID 19.

- 01. The machine should be turbine / compressor based because the installation sites might not have central Oxygen Lines.
- 02. The machine should have Invasive, Non- Invasive and CPAP features to make them versatile.
- 03. 200-600 L tidal volume, Lung Mechanics Display
- 04. Monitoring of plateau Pressure, PEEP, PS, Oxygen Concentration, Lung mechanics /inverse ratio(I:E)
- 05. Pressure & Volume control & PSV
- 06. Continuous working capability for 4-5 days
- 07. FiO2 manipulation b/w21-100% is essential
- 08. PEEP upto 30
- 09. FiO2 delivery should be independent of the flow rate PEEP and the minute volume

 FiO2 delivery to the patient can't be made dependant on other factors. Every patient has a different lungs dynamics and can't be delivered gases at the same settings.

Senior Manager
Sourcing Division - RBD
HLL Lifecare Ltd.
HLL Bhavan, Poojappura,
Thiruvananthapuram.
Ph.no: 0471 2353932.