

**Pre bid meeting minutes held at HLL Biomedical office on 18.8.2014 at 11.15 am for the Supply of Medical Furniture and Medical Equipment for the Strengthening & Upgradation of Govt: Medical College ,Kozhikode**

**Tender No: HLL/ID/14/49**

**SI No.4. Crashcart**

<b><u>SN</u></b>	<b><u>Existing specification</u></b>	<b><u>Amended Specification</u></b>
	With five lockable drawers of extruded aluminium or MS powder coated to hold emergency drugs.	With six removable bins and two polystyrene lockable storage unit to hold emergency drugs.
	Body of metal sheets and frame of 20mm square MS tubes. Metal parts pretreated and epoxy powder coated. Top of SS/ ABS with assorted surgical instruments for resuscitation emergency care,	Should be with S.S 304 tubular frame and SS 304 shelves. Top of SS 304 /ABS with provision for mounting medical equipment & assorted surgical instruments for resuscitation emergency care.
	NIL	Overall dimension : 960 mm X 500 mm X 1500approx

**SI No.5 Fully Automated Rotary Microtom**

<b><u>SN</u></b>	<b><u>Existing specification</u></b>	<b><u>Amended Specification</u></b>
1	Cold light Source	Deleted
2	Motorized coarse feed in two speeds 30 micron/sec and 90 micron/sec. Variable sectioning speed adjustable from 0.5 to 420 mm/sec.	Motorized coarse feed in two speeds 300 micron/sec and 900 micron/sec. Variable sectioning speed adjustable from 0.5 to 420 mm/sec.
3	Disposable blade holder with lateral displacement feature that can hold both high and low profile blades and knife holder which can accommodate 16 cm C& D type knives.	Disposable blade holder with lateral displacement feature that can hold both high and low profile blades and optional knife holder which can accommodate 16 cm C& D type knives.
4	C Type knife 16 cm - 3Nos	C Type knife 16 cm - 3Nos-Optional (Price should be quoted separately in price bid)
5	Cold plate ( Dry type)	Cold plate ( Dry type)-Optional (Price should be quoted separately in price bid)

**SI No: 7. Fowler Cot and bed for Various Department**

<b><u>SN</u></b>	<b><u>Existing specification</u></b>	<b><u>Amended Specification</u></b>
1	The quoted model should have CE/BIS certificate and copy of the same should be enclosed along with the technical bid.	deleted
2	NIL	Should be provided with a SS hook at suitable position to hang the Urine bag.

The tender submission date and opening date is also amended as follows:

<b>Tender No:</b>	<b>Name of the work</b>	<b>Tender submission date and time</b>	<b>Tender opening date and time</b>
HLL/ID/14/49	Supply of Medical Furniture and Medical Equipment for the Strengthening & Upgradation of Govt: Medical College ,Kozhikode	27.08.2014 at 11.00 am	27.08.2014 at 12.00 pm

The form 16 will be issued by the **Consignee, Govt: Medical College ,Kozhikode**, on production of details by the suppliers duly verified by the Engineer in charge.

The pre bid meeting minutes shall form part of the tender document and all the bidders are requested to seal and sign the pre bid meeting minutes and should be submitted along with the tender.