

**Amendment No. 3**

Date: 25/07/2014

**Subject: Amendment to the Tender Enquiry Document**

**Ref: NIT ref.: HLL/PCD/GNCTD/07/MAMC/14-15 dated 04/06/2014**  
**Tender ID in GNCTD e-Procurement portal: 2014\_HFWD\_58635\_5 &**  
**2014\_HFWD\_58635\_6**

The following changes have been incorporated in the referred tender.

**SECTION I**  
**NOTICE INVITING TENDERS (NIT)**

**Existing:**

Sl. No.	Tender_ID	Description	Qty.	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	Closing date & time for submission of physical Tender	Date & time of opening of tender
5	2014_HFWD_58635_5	Laparoscopic Simulator Portable	2	3,000	400,000	10-06-2014 11:00 AM	30-07-2014 06:00 PM	31-07-2014 01:00 PM	31-07-2014 02:00 PM
6	2014_HFWD_58635_6	Laparoscopic Wet Skills Station (consisting of Laparoscopic set complete)	10	3,000	400,000	10-06-2014 11:00 AM	30-07-2014 06:00 PM	31-07-2014 01:00 PM	31-07-2014 02:00 PM

**Read as:**

Sl. No.	Tender_ID	Description	Qty.	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	Closing date & time for submission of physical Tender	Date & time of opening of tender
5	2014_HFWD_58635_5	Laparoscopic Simulator Portable	2	3,000	400,000	10-06-2014 11:00 AM	<b>07-08-2014</b> 06:00 PM	<b>08-08-2014</b> 01:00 PM	<b>08-08-2014</b> 02:00 PM
6	2014_HFWD_58635_6	Laparoscopic Wet Skills Station (consisting of Laparoscopic set complete)	10	3,000	400,000	10-06-2014 11:00 AM	<b>07-08-2014</b> 06:00 PM	<b>08-08-2014</b> 01:00 PM	<b>08-08-2014</b> 02:00 PM

**Section – VII**  
**Technical Specification**

**Item no. 5-****Laprosopic Wet Skill Station****Existing:****Digital Camera (Camera Head and Console)**

**Read as:**

**Added Para: A full high definition camera should be provided.**

**Existing:**

4. Medical Grade Monitor

Coloured Monitor 26" Flat Monitor, PAL system compatible. Composite, S-Video and DVI inputs, Compact & Lightweight design, the monitor should support Direct Fibre input. Should be US FDA or European CE approved.

**Read as:**

4. Monitor

At least 26" full HD monitor compatible with Camera.

**Existing:**

2. Xenon Light Source

220 Volts, 300 watts xenon bulb with elliptical bulb design, high colour temperature- more than 6000K corresponds to brightness of sunlight resulting in high visual and photographic clarity for colour retention, Monitoring of lamp function. Bulb Life Counter on Light source, standby mode, and universal jaw assembly to adapt any make of fibre optic cable, light intensity adjustment continuously adjustable from 0 to 100% manually with spare bulbs to give additional 5000 hours of added running life- 10 Nos. Should be US FDA or European CE approved.

**Read as:**

2. Light source

LED light source (175 W-300 W) with one spare LED bulb

**Existing:**

Mobile Video Cart

**Read as:**

Mobile cart

- a) For Laproscopic stack- Mobile high quality cart with multiple height adjustable shelves with atleast one drawer with 4 high quality castor wheels (2 lockable) with built in atleast 6-8 multi power sockets (5/15 ampere) with single main power switch for entire stack.
- b) For Endo Trainer Box- Mobile high quality cart with two height adjustable shelf with at least one drawer (with 4 high quality castor wheels (2 lockable)).  
Both Indian makes acceptable.

**Added Para:**

1. **Electronic UPS (atleast 1 KVA) to provide back up for all the laproscopic power supply items for 10-15 minutes.**
2. **Instruments:**  
**The following instruments to be added (all reusable):**
  - a. **Maryland-2**
  - b. **Grasper-2**
  - c. **Metzenbaum Scissors-1**

- d. Hook Scissors-1
- e. Driver Needle holder-1
- f. Assistant needle Holder-1

**All items except mobile video cart and endotrainer to be manufactured by the same company**

**Item no. 6**  
**Laposcopic Simulator Portable**

**Existing:**

5. Provide various operative procedure modules like Laproscopic cholecotomy, fundoplication incision hernia repair inguinal hernia repair, colorectal surgery and bariatric surgery.

**Read as:**

5. Provide various operative procedure modules like Laproscopic cholecotomy, **Laposcopic Appendicetomy, Laproscopic Hysterectomy, Laproscopic Nephrectomy**, fundoplication incision hernia repair inguinal hernia repair, colorectal surgery and bariatric surgery.

**All other contents of the tender enquiry including terms & conditions remain unaltered.**

**Note: Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids**