

AMENDMENT NO.1 Dated 10.June.2017

Ref Tender No: HLL/SD/CHO/2017-18/TENDER/002

The following amendments have been incorporated to the bid document for the above Tender for Supply of Diagnostic Equipments under Outright Purchase to HLL Lifecare Limited's diagnostic Laboratories across India.

Sl No	Parameters			
	Original Tender Detail			
	Annexure 1 <u>Equipments required</u>			
	SL.NO	EQUIPMENT	ESTIMATED QUANTITY (IN NOS)	EMD
	1	SEMI AUTO ELISA READER	2 Numbers (Indicative)	Rs.6,000/- (Rupees six thousand only)
	2	SEMI AUTO BIOCHEMISTRY ANALYSER	57 Numbers (Indicative)	Rs.3,00,000/- (Rupees Three Lakhs only)
	Amended tender Details			
	ANNEXURE – 1 <u>Equipments required</u>			
	SL.NO	EQUIPMENT	ESTIMATED QUANTITY (IN NOS)	EMD
	1	SEMI AUTO ELISA READER	4 Numbers (Indicative)	Rs.16,000/- (Rupees sixteen thousand only)
	2	SEMI AUTO BIOCHEMISTRY ANALYSER	65 Numbers (Indicative)	Rs.2,00,000/- (Rupees Two Lakhs only)

AMENDMENT NO.1 Dated 10.June.2017

Amended tender Details

ANNEXURE 1.1 Specification of SEMI Auto Elisa Reader shall be read as Below:

SI. No.	SEMI AUTO ELISA READER	Whether the product meets the technical spec (Yes/No)
1	Should be 8-channel optical measuring system.	
2	Should work with a keypad or touch screen / PC System .	
3	Should be able to read 96-well plate.	
4	Should have capability for Bi chromatic measurements.	
5	Measurement time 12 seconds or less in the Monochromatic measurement mode.	
6	Should be provided with standard 405nm, 450nm, 492/490 nm, 570/578 nm, 630nm and 700 nm standard filters.	
7	Should have large LCD display/ attached to PC, with user friendly software operation.	
8	Should have photometric range 0.0 ~ 3.5 or more Abs	
9	The On-board software should have capability of storing the calibration curve data for at least 8 standards in all the test programs.	
10	Should have memory of at least 2400 sample test results	
11	The system must accept external printer and must print results in pre-formatted matrix form giving details such as Sample No., Value, Abs and interpretation, with cut-off equation for qualitative results.	
12	Should have ports for external printer and for transmission of data to the host computer.	
13	Necessary Online UPS with One hour back up should be provided.	

AMENDMENT NO.1 Dated 10.June.2017

14	The company supplying the instrument should have service and distributor network all over India	
15	Unconditional application support must be provided for programming any third party kit in the machine.	
16	The supplier has to provide 3 Years unconditional warranty	

Amended tender Details

A NNEXURE 1.2 , Specification of SEMI Auto Biochemistry Analyser shall be read as Below:

Sl. No.	SEMI AUTO BIOCHEMISTRY ANALYSER	Whether the product meets the technical spec (Yes/No)
1	Analyzer should be microcontroller based semi-automated analyzer to perform routine biochemistry tests in Monochromatic/Bichromatic Mode (endpoint, fixed time & kinetic chemistries, multi standard curve calibration etc)	
2	Analyzer should offer minimum of 100 user definable programmable parameters.	
3	Analyzer should have Minimum 6 static interference filters	
4	Analyzer should display real time graph with averaged reading for each second from start to finish of the test.	
5	Analyzer should have aspiration volume of 350 µl or less	
6	Analyzer should have facility to store around 1000 test results in the memory and can recall stored test results in ID wise, Date wise, Date and Test wise etc.	
7	Analyzer should have levy Jennings control graphs and Westgard algorithm	
8	Analyzer should have flow cell with volume capacity of 30-36 µl or less and temperature selection settings programmable from 25°C - 37°C .	

AMENDMENT NO.1 Dated 10.June.2017

All other terms and conditions in the original Tender Notice and tender documents remain same and unchanged.

For HLL Lifecare Ltd.

Senior Manager (Sourcing Division)