

Amendment No. 1

Date: 15/03/2014

Subject: Amendment to the Tender Enquiry Document**Ref: NIT ref.: HLL/PCD/GNCTD/02/13-14 dated 12/02/2014
Tender ID in GNCTD e-Procurement portal: 2014_HFWD_52736**

The following changes have been incorporated in the referred tender .

SECTION I**Existing:**

Description	Total Qty.	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	Closing date & time for submission of physical Tender	Date & time of opening of tender
C-Arm Compatible Paeditrics OT Table	2	2,000	170,000	20-02-2014 11:00 PM	17-03-2014 6:00 PM	18-03-2014 1:00 PM	18-03-2014 2:00 PM

Amended as:

Description	Total Qty.	Tender Fees (Rs.)	EMD Amount (Rs.)	Date & time of Prebid meeting	Date & time of closing of online tender	Closing date & time for submission of physical Tender	Date & time of opening of tender
C-Arm Compatible Paeditrics OT Table	2	2,000	170,000	20-02-2014 11:00 PM	02-04-2014 6:00 PM	03-04-2014 1:00 PM	03-04-2014 2:00 PM

Section – VII
Technical Specifications

C-Arm Compatible O.T. Tables for Children

Sl. No	Tender Page No.	Existing Specification	Read As
1	Page No: 43 Section – VII Technical Specification	2. It should have mobile battery and power operated electromechanical table with manual override (hand and foot).	2. Table should be electro-hydraulic with battery backup with manual override (hand/foot)
2	Page No: 43 Section – VII Technical Specification	16. The electro-mechanical adjustment for smooth and precise operation of head, upper back, lower back, leg section and kidney and all surgical and urological procedures should be possible.	16. The electro-hydraulic adjustment for smooth and precise operation of head, upper back, lower back, leg section and kidney and all surgical and urological procedures should be possible.
3	Page No: 43 Section – VII Technical Specification	18. The table should be manufactured by a reputed company having ISO9001 / BIS standards.	18. The table should be USFDA and/or European CE approved.

All other contents of the tender enquiry including terms & conditions remain unaltered.