

बिड दस्तावेज़ / Bid Document

बिड विवरण / Bid Details	
बिड बंद होने की तारीख/समय / Bid End Date/Time	12-06-2025 15:00:00
बिड खुलने की तारीख/समय / Bid Opening Date/Time	12-06-2025 15:30:00
बिड पेशकश वैधता (बंद होने की तारीख से) / Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम / Ministry/State Name	Ministry Of Health And Family Welfare
विभाग का नाम / Department Name	Department Of Health And Family Welfare
संगठन का नाम / Organisation Name	Hll Lifecare Limited
कार्यालय का नाम / Office Name	Hll Bhavan, Registered Office, Poojapura
क्रेता ईमेल / Buyer Email	buycon47.hll.kl@gembuyer.in
कुल मात्रा / Total Quantity	1
वस्तु श्रेणी / Item Category	Advanced Integrated Video Urodynamic Workstation (Q3)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) / Minimum Average Annual Turnover of the bidder (For 3 Years)	184 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष / Years of Past Experience Required for same/similar service	3 Year (s)
टर्नओवर के लिए एमएसई को छूट प्राप्त है / MSE Exemption for Turnover	Yes
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है / Startup Exemption for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़ / Document required from seller	Experience Criteria, Past Performance, Bidder Turnover, Certificate (Requested in ATC), OEM Authorization Certificate, Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer

बिड विवरण/Bid Details	
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	No
विगत प्रदर्शन /Past Performance	30 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	Yes
रिवर्स नीलामी योग्यता नियम/RA Qualification Rule	50% Lowest Priced Technically Qualified Bidders
बिड का प्रकार/Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
मूल्यांकन पद्धति/Evaluation Method	Total value wise evaluation
वित्तीय दस्तावेज की आवश्यकता है / Financial Document Required	Yes
मध्यस्थता खंड/Arbitration Clause	No
सुलह खंड/Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	HDFC Bank
ईएमडी राशि/EMD Amount	736000

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	62

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई कैटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

DGM-CSR

Power Grid Corporation of India Limited Northern Region - I HQ, SCO Bay No.5 to 10, Sector 16A, Faridabad - 121 002, Haryana

(Dgm -csr)

विभाजन/Splitting

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता/MII Purchase Preference

एमआईआई खरीद वरीयता/MII Purchase Preference	Yes
--	-----

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
---	-----

1. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be exempted from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking exemption from Turnover, shall upload the supporting documents to prove his eligibility for exemption.
2. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
3. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
4. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

[OM No.1_4_2021_PPD_dated_18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy

for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

5. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

6. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

7. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 30% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

8. Short Duration Bid has been published by the Buyer with the approval of the Competent authority due to Emergency procurement of critical products/services.

9. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available. Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

एक्सेल में अपलोड किए जाने की आवश्यकता /Excel Upload Required :

BOQ29 - [1749114333.xlsx](#)

Advanced Integrated Video Urodynamic Workstation (1 pieces)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Technical Parameters	Integrated Air pump for UPP/ARM	Yes, No
	Technology used for Data Prevention	NDL Technology
	Ambulatory Urodynamics	Reusable Catheter with live streaming facility, Should have leak Point Detection Pad, Option for upgradation in future for UPP and Anorectal Manometry, It should have Uroflow Transducer, Laptop and Urodynamic Software
	Type of EMG	It should have both visual as well as sound EMG
	Leak Detection Module	Leak Detector should be placed at end of the urethra to detect even a small drop of leak which do not come into beaker and should automatically mark leak point event in the software during the study
	EMG Settings	It should have adjustments for gain, threshold and Volume
	Nomograms	ICS Nomogram, Siroky, Purr, Schaffer Nomograms, Paediatric nomogram
	Remote Control	Yes, No
	Video Urodynamic Software	Provision for recording live videos during the procedure of Urodynamic study, Digital Video Urodynamic Software for doing Video Urodynamic studies, Facility to take image and video with the help of CARM and store in File, It should have facility for Synchronization with Urodynamic Traces
	Wireless EMG	Yes, No
	Leak Point Pressure Recording	Yes, No
C-Arm	C-ARM Adjustments	Supine to 90 degree Standing Position, Seated to 90 degree Standing Position, Supine to Seated Position
	Accessories	Hand Control, Two Foot Rests, One fixed patient ARM Rest, One Plastic Funnel drainage Detachable arm rests, Armrest holders, One detachable side rail
	C-ARM compatible Three Motors Video Urodynamic and Fluoroscopy Table Features	Facility of standing, seated and supine positioning capabilities via hand held control, Should be compatible with video urodynamics studie, Should be side-mounted cantilevered design offers the physician anterior or posterior, lateral and oblique views without obstruction, Should have the c-arm access and a wide range of positioning possibilities, make it ideal table for any Fluoroscopy Procedure, Should have three motors which offers unparalleled positioning or access options, Should have radiolucent carbon fiber table top for head-to-toe Imaging, Should have Double wheel rolling casters offer unencumbered Maneuverability
	C-ARM Compatible Three Motors Video Urodynamic and Fluoroscopy Table	Yes, No

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
Near Infrared Spectroscopy (NIRS)	It should be supplied with Portable One Channel of Near Infrared Spectroscopy (NIRS) to understand and measure changes in Oxy, Deoxy and total haemoglobin and tissue saturation.of bladder muscle to correlate with Urodynamic studies	Yes, No
	Technology used	Continuous Wave Near Infrared Spectroscopy
Generic	Salient Features	Integrated Travel Cart, All in One PC, Rust Free Trolley, Control Panel Completely inside to avoid water Spillage, Integrated water proof keyboard, Touch pad, Mouse, Highly Sensitive Radio Frequency Based weight transducer, Wireless EMG having Audio and Visual with adjustment of threshold, Sound, Gain, 3 motors fluroscopy Video Urodynamic procedure table cum chair for sitting, Supine and Standing position for doing Urodynamic studies, Leak Point Detector to detect the even small drop of leakage which does not come in Uroflow, Integrated infusion sensor to double check the infused volume

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values)
	Inclusion	Advanced Integrated Video Urodynamic Workstation for Urology and gynaecological investigation for Adult & Pediatric patients as per above specification, The Computer running on Window 10, i7 processor, 1 TB HDD, 8 GB RAM, Integrated Waterproof Keyboard with Touchpad, CD/DVD writer, USB Ports, 20-24 inch Color LED monitor, Video editing software alongwith latest office professional edition, Optical Mouse, Speakers for EMG and color laser printer and UPS of minimum 30 min. back up, Urodynamic Pressure Kit, Wired RF Based Uroflow Transducer, Wireless RF Based Uroflow Transducer, Wireless EMG with Visual and Sound, Leak point detector module to detect even a small drop of leakage which do not come into Uroflow, Height Adjustable Folding Micturation Chair with detachable arm for Urodynamic studies with removable funnel, Height Adjustable Uroflow stand with removable funnel, Digital Video Urodynamic Software and Hardware for doing Video Urodynamic Studies, EMG Biofeedback Software for complete pelvic floor assessment and training with international standard games with Anal Probe and Vaginal Probe for Biofeedback, 4 Channel Anorectal Manometry software for doing Anorectal Manometry, C-ARM compatible Three Motors Video Urodynamic & fluoroscopy table, Ambulatory Urodynamic complete set with reusable catheter with live streaming facility, Near Infrared Spectroscopy (NIRS) for dynamic topography of pelvic muscle to understand the hydrogenation and dehydrogenation of Pelvic/Bladder muscle to correlate with Urodynamic Studies, 2 Lumen Catheter for Cystometry 8 Fr, 2 Lumen Catheter for Cystometry 6 Fr, 2 Lumen Pigtail Catheter for Cystometry 5 Fr, 3 Lumen Catheter for UPP, Rectal Balloon Catheter 10 -12 Fr, 4 Channel ARM Catheter longitudinal with balloon (single use), 4 Channel ARM Catheter longitudinal with balloon (reusable), Dome for pressure transducer - (Disposables), Pump Tube, Pressure Cuff for Perfusion, Urethra Perfusion line, Surface electrodes for EMG, Connection tubings, Transparent Security Cover for Uroflow transducer, Anal Probe for biofeedback, Vaginal Probe for biofeedback, Reusable Bladder catheters for Ambulatory urodynamic, Reusable Rectal catheters for Ambulatory urodynamic, Leak detection pad for ambulatory urodynamic
	Services provided under Warranty	Urodynamic System, 3 motors fluroscopy Video Urodynamic procedure table cum chair, Ambulatory Urodynamics, Bladder Scanner, Near Infrared Spectroscopy System} Spare Parts, Accessories, Labour Charges
	Warranty	1, 2, 3, 4, 5 (year)

Additional Specification Parameters - Advanced Integrated Video Urodynamic Workstation (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
TECHNICAL SPECIFICATION	The specification provided in GEM tender document is only for reference and for actual specifications Bidders are requested to refer the Tender Document uploaded by HLL as ATC under the heading 'Buyer Added Bid Specific Terms and Conditions'.
PRODUCT COMPLIANCE SHEET	Bidder has to provide item by item compliance sheet as per attached ATC specification

* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी के दिन/Delivery Days
1	Prasad K R	695012,HLL LIFECARE LTD, HLL BHAVAN, POOJAPURA, THIRUVANANTHAPURAM	1	45

Special terms and conditions-Version:1 effective from 18-11-2022 for category Advanced Integrated Video Urodynamic Workstation

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation(CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
4. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
5. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address,telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled.Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
6. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
7. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any),

rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.

8. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.

OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.

9. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
10. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
11. **Software:** All software updates should be provided free of cost during warranty period.
12. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Generic

OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 50 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 50% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

2. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

Technical Specifications of Advanced Integrated Video Urodynamic System

1	<p>The unit should be able to perform CMG, Pressure flow, EMG, EMG with biofeedback, Video Urodynamics with compatible C ARM, Uroflowmetry (with one wired and one wireless uroflowmeter), Urethral Pressure Profilometry, Anorectal Manometry, Ambulatory Urodynamics & fNIRS for Bladder & Brain.</p> <p>a) System should be supplied with EMG having both visual & sound EMG (both needle & patch EMG)</p> <p>b) Should be supplied with a separate wireless EMG with biofeedback software/Module configured for pelvic floor training (having animated games for patient engagement). Should be supplied with Anal Probes (5 Nos) and Vaginal Probes (5 Nos)</p>
2	<p>Video Urodynamic Chair:</p> <ol style="list-style-type: none"> Radiolucent motorised chair providing eight positioning functions capable of supine, sitting and standing position It should allow patient to remain in one place for preparation, filling and voiding phases of the urodynamic test It should eliminate the risk of catheters falling out or getting misplaced during transfer or ambulation Chair should be protected against water incursion There should be double wheel rolling casters offering unencumbered maneuverability There should be side mounted cantilevered design offering anterior/posterior, lateral and oblique views without obstruction and to provide easier access to, and visualization of, the perineal area. There should be radiolucent carbon fiber seat and back for head-to-toe imaging. Weight capacity in standing position should be 250kg.
3	Equipment should have 3 or more pressure channels.
4	Should display atleast 6 channels at a time
5	Software for multiple Calibrations and multiple connections.
6	Transducer should be reusable with automatic zero facility, pressure range of (-40 - 350cm of H ₂ O).
7	Should be supplied with Two weight Based Uroflow Transducers, One wired and another wireless with latest RF/Xbee/Jigbee Technology with flow range of 0-50 ml/sec and volume range upto 1500 ml. Should have auto record and zeroing facility for uroflowmeter and auto artefact recognition and compensation. Should be supplied with Protecting cover for flow transducer, uroflow stand and 2 litre beaker.
8	Should have running Tap water Sound and should have No Data Loss Technology in software.
9	Should have infusion volume upto 1000ml & software based calibration control and have the ability to show infusion volume

10	Should be supplied with integrated travel cart to fit the complete UDS along with DC/PC.
11	Should be supplied with All in one PC computer running on Windows 10 or more, i7 processor, at least 16 GB RAM, 1 TB SSD Memory, DVD/CD rewritable drive (external) 1GB graphic card, Microphone, speaker for EMG, Video Editing software, 22-26 inch flat colour LED screen monitor, MS office, USB port, Integrated waterproof keyboard, Optical mouse with mouse/touch pad, UPS & color Ink tank printer
12	Provision for recording live videos during the procedure of urodynamic study. The system should have ability to show static images with corresponding Urodynamic Trace at that time point
13	Should be supplied 21" or more flat screen monitor for display. Monitor should have facility of RGB output so that it can be projected on screen in conferences
14	Automatic Leak detector module to detect even small drops of urine leakage which do not come into beaker. The Leak detected should be automatically marked in the software.
15	Should have wireless data transfer facility for CMG, Pressure flow and other Urodynamic studies.
16	Should have ICS nomogram, Siroky, Purr, Schaffer Nomograms and Paediatric Nomograms inbuilt in the software.
17	Should have inbuilt pump for infusion with filling rate of 1 ml/min - 100 ml/min with software controlled pump calibration with filling volume of 0-1500 ml.
18	<p>Should be supplied with following consumables (Catheters should be from the same manufacturer):- (Unit price of all consumables should be quoted separately for future purchases)</p> <ul style="list-style-type: none"> 2 Lumen Catheter 8 Fr. For Cystometry - (Qty 70 pcs) 2 Lumen Catheter 6 Fr. For Cystometry- (Qty 30 pcs) 2 Lumen 5 Fr Pigtail Catheter for Cystometry - (Qty 20 pcs) Rectal Catheter should be between 10-12 fr. -(Qty 120 pcs) Pump Tube - (Qty 50 pcs) Surface electrodes for EMG - (Qty 360 pcs) Reusable pressure transducers - (Qty 10 pcs) Disposable Domes- (Qty 50 pcs) Connection tubings - (Qty 240 pcs) Collection Jug 2Ltr - (Qty 5 Nos) Infusion sets - (Qty 50 Nos) Protective Plastic cover for Flow transducer - (Qty 25 Nos) UPP Catheter 3 Lumen - (Qty 10 Nos) ARM Catheter 4 Channel 1cm long with balloon - (Qty 10 Nos) 500 Gm Calibration weight - (Qty: 1 No) Removal funnel - (Qty 3 Nos) Ultrasound Gel - (Qty:20 Bottles) Thermal Printer Paper Roll - (Qty 10 pcs) A4 sheet Paper - (Qty 10 pack) Batteries for Uroflow - (Qty 16 Nos) Color Printer Cartridge - (Qty 4 Nos)

19	Should be supplied with UPP Puller and Software for doing urethral Pressure Profilometry.
20	Should be Supplied with 4 Channel Anorectal Manometry hardware and software with catheters for doing Anorectal Manometry.
21	Should be supplied with Ambulatory Urodynamics Complete set with catheters and Leak detection Pad having live streaming facility. Should be supplied with Laptop, Dongle, software with key, event marker, transducers and catheters for initial 100 cases.
22	Should be supplied with FNIRS (Functional Near Infrared spectroscopy) two numbers. One for Bladder muscle and another for Brain.
23	C Arm with Dynamic Flat Panel detector compatible with the UDS machine for Video Urodynamic Studies should be provided. The quoted model should be USFDA / European CE with 4 digit notified body / BIS approved. Should be supplied with Thyroid Shield -5nos, Lead Apron -5Nos & Stand -1 no.
24	Should be supplied with 3D Bladder scanner for PVR measurement for complete urodynamic studies
25	<p>The bidders are required to modify Video Urodynamic Lab of Area around 400 sqft to make it convenient to run Video urodynamic study with C-ARM. This includes following</p> <ul style="list-style-type: none"> A. Led Protection on Doors & Windows as per AERB guidelines B. Internal partitions for Patient preparation and procedure and faculty C. One Plasma TV for Biofeedback (42 inches or more) D. Wooden work like cabinet, cupboard, Almirah etc E. Water Drainage , Basin etc F. Internal Electrical requirements like FAN, Tubes, A/C, Wiring , Cables etc in Lab G. Painting, Plaster and Designing of walls and Tiles on floor
26	The quoted equipment must be approved by US FDA / European CE Certified with four digit notified body / BIS certified.
27	Demonstration whenever required should be arranged. The firms must have installation of quoted / similar model.
28	Five years of comprehensive warranty followed by five years of comprehensive annual maintenance contract (CAMC). Comprehensive warranty as well as CAMC shall include all parts. CAMC rates should be quoted in price bid. Warranty and CAMC on rates quoted by local agent for all items must be guaranteed by principal/OEM via an undertaking.
29	Rates of consumables, spare parts and additional accessory items to be quoted separately by the firm. These quoted rates should be valid for at least 3 years. Anything which is not separately mentioned as a consumable will be considered to be a spare part and which would then be covered under warranty / CAMC
30	There will be 95% uptime warranty during CMC period on 24(hrs) x 7(days) x 365 basis, with penalty, to extend CMC period by double the downtime period. If non functional period exceeds two weeks at a stretch, a temporary standby machine of same specification should be provided.

3. Buyer Added Bid Specific ATC

Buyer uploaded ATC document [Click here to view the file.](#)

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976,

The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---