

TENDER ENQUIRY NOTICE

HLL/ID/Ktym-Med Equip/2014-15 /

21.11.2014

Name of Work : Supply of 2 no:s of Pharmaceutical Refrigerator for the setting up of emergency and critical care department at Kottayam Medical College.

Limited tenders are invited on behalf of Principal, Kottayam Medical College, Govt. of Kerala for the Supply of 2 no:s of Pharmaceutical Refrigerator for the setting up of emergency and critical care department at Kottayam Medical College. The specification for the Pharmaceutical Refrigerator is enclosed herewith.

Address to which the offers are to be submitted:

Deputy Vice President (Technical)

HLL Lifecare Limited

“ADARSH”, TC - 6/1718 (1)

Vettamukku, Thirumala P O

Thiruvananthapuram - 695 006

Ph : 0471 - 2365872/73

Fax: 0471 - 2368144

Submission of Offer

The offer shall be submitted in single bid system. The offer should be submitted in a sealed envelope superscribing “Supply of 2 no:s of Pharmaceutical Refrigerator for the setting up of emergency and critical care department at Kottayam Medical College.”

Last date and time of Receipt of offers : 29.11.2014 11.00 am.

Date and time of Opening of Offers : 29.11.2014 11.30 am

Payment terms

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

a) On delivery:

75 % payment of the contract price shall be paid on receipt of goods in good condition and upon the submission of the following documents:

- (i) Four copies of supplier’s invoice showing contract number, goods description, quantity, unit price and total amount;

Other Conditions

1. Warranty should be for minimum 2 years from the date of acceptance of Pharmaceutical Refrigerator and the Warranty has to be given in the name of client.

Performance Security

2. Within Fifteen (15) days from date of the issue of purchase order the successful bidder, shall furnish performance security for an amount equal to ten percent (10%) of the total value of the contract, valid up to sixty (60) days after the date of completion of all contractual obligations by the supplier, including the warranty obligations.
3. Demo of equipments and sample of instruments should be arranged when asked for
4. Earnest Money of Rs. 12,000/- has to be submitted along with the tender in the form of a Demand Draft/ Fixed Deposit Receipt (FDR) of a scheduled bank issued in favour of HLL Lifecare Limited, Thiruvananthapuram which should be placed in a separate sealed cover marked "Earnest Money" shall be submitted along with the tender.
5. The rate quoted should be inclusive of all taxes, duties, transport, loading and unloading up to consignee site.
6. The tenderer shall submit the compliance for technical specifications. The make and Model should also be attached along with the bid.
7. The original Literature or the relevant part of the user/service manual should be attached as proof.
8. The offer shall be valid for 3 months.
9. The invoice/ bills shall be raised in the name of "**Principal, Kottayam Medical College a/c HLL Lifecare Ltd**" and the same shall be submitted to the HLL Engineer in charge.
10. TIN no. of HLL should not be mentioned in any of the documents as HLL is acting on behalf of the client in the capacity of consultant alone.
11. The completed works shall be handed over to the client after due verification by HLL Engineer in Charge.
12. The bidders should seal and sign the Tender Enquiry Notice and should be submitted along with the offer.

Deputy Vice President (Tech)

HLL Lifecare Limited

PHARMACEUTICAL REFRIGERATOR	
1	Description
	Used for storing blood plasma and other blood products, vaccines, other medical or pharmaceutical supplies. Also to cool samples or specimens for preservation
2	Technical Specifications
2.1	Capacity: 300 - 400 L
2.2	Temperature : Temperature 2-8° C.
2.3	Material of Construction: SS 304 interior and outer with powder coated CRCA.
2.4	Insulation: High-grade pressure - foam material.
2.5	Door: Lockable door with plastic magnetic sealing surround.
2.6	Air-Cooling System : Should have re-circulating air-cooling system.
2.7	Control Panel Features : Control panel with temperature alarm, ON /OFF switch with power on indicator, digital display of temperature
2.8	Compressor Type : Hermetically enclosed, low noise, vibration proof compressor.
2.9	Alarm : Audio visual alarm system.
2.10	Other Features
	Should be CFC free & low noise
	Should have automatic defrosting and condensed melt water evaporation.
	No welded joint to be exposed for rusting.
	Should have Interior lighting & Adequate circulation of air to ensure even cooling
3	Standards, Safety and Training
3.1	Should have the ISO certification and the copy of the same should be enclosed along with the technical bid.
4	Documentation
4.1	User/Technical/Maintenance manual to be supplied in English
4.2	Certificate of calibration and inspection from factory.
4.3	Warranty & CMC as per tender terms.
4.4	Must submit at least 3 user list and performance report within last 5 years from major hospitals.
4.5	Must submit at least 2 nos of latest purchase order of the quoted model dated within 1 year along with the price bid .