

29-02-2020

Amendment No 04**Sub: Amendment to the referred tender enquiry****Ref.: Tender Enquiry Document HITES/PCD/PMSSY-IV/07/RAD/19-20 dated 16-12-2019, Amendment No 01 dated 15-01-2020, Amendment No 02 dated 27-01-2020, Notification 01 dated 30-01-2020, Amendment No 03 dated 15-02-2020 and Notification 02 dated 28-02-2020.****SECTION I****NOTICE INVITING TENDER (NIT)****For Tender IDs 2019 HLL 38075 1 and 2019 HLL 38075 2 only:-****Tender Timeline**

Sl. No.	Description	Schedule
1	Closing date & time for submission of online bids	03-03-2020, 01:00 PM
2	Last date and time of physical submission of EMD, Tender Processing Fee, any other document specified in the Bidding Document	04-03-2020, 02:00 PM
3	Time and date of opening of online bids	04-03-2020, 02:30 PM

Read As:

Sl. No.	Description	Schedule
1	Closing date & time for submission of online bids	17-03-2020, 01:00 PM
2	Last date and time of physical submission of EMD, Tender Processing Fee, any other document specified in the Bidding Document	18-03-2020, 02:00 PM
3	Time and date of opening of online bids	18-03-2020, 02:30 PM

Tender Timeline for all other Tender IDs remain unchanged.

Section VII**Technical Specification**

Tender Page & Para	CURRENT TENDER SPECIFICATION	READ AS
Item No 01 - 1000 mA Digital X-RAY System with Digital Flat Panel Detector		
Pg 44 Para 2.e	Grey scale of 14 bit or more	Grey scale of 12 bit or more.
Pg 46 Para 9.f	Automatic stitching should be available in table as well as in chest stand. Image acquisition software & stitching software should be from the same Principal manufacturer and should be US FDA/European CE/BIS approved	Automatic stitching should be available in table as well as in chest stand. Image acquisition software & stitching software should be from the same manufacturer /Principal manufacturer and should be US FDA/European CE/BIS approved

Tender Page & Para	CURRENT TENDER SPECIFICATION	READ AS
Item No 02 DRF - Digital Flat Panel Fluoroscopy cum Radiography System		
Pg 52 Para D.1	Field of view of at least 43 x 43 cm or more / (35 x 43 cm with rotating facility)	Field of view of at least 43 x 43 cm or more
Pg 52 Para E.2	Detector must be at least 43 x 43 cm or more / (35 x 43 cm with rotating facility)	Detector must be at least 43 x 43 cm or more

All other contents of the tender enquiry including terms & conditions remain unaltered

Note:

- i. Only for following Tender IDs Techno Commercial Sheet v1.2 is being uploaded on CPP Portal. The same need to be downloaded for submission. It will supersede any other version uploaded earlier.**
 - a. 2019_HLL_38075_1 - 1000 mA Digital X-RAY System with Digital Flat Panel Detector.**
 - b. 2019_HLL_38075_2 - DRF - Digital Flat Panel Fluoroscopy cum Radiography System.**
- ii. For following Tender ID Techno Commercial Sheet v1.2 was uploaded on CPP Portal. The same need to be downloaded for submission.**
 - a. 2019_HLL_38075_13 - Picture Archiving & Communication System (PACS).**
- iii. For all other Tender IDs, Techno Commercial Sheet v1.1 was uploaded on CPP Portal. The same need to be downloaded for submission.**
- iv. Price Bid format v1.1 for this tender was uploaded on CPP Portal. The same need to be downloaded for submission.**
- v. Prospective Bidders are also advised to check the website regularly prior to the closing date and time of online submission of bids.**