

AMENDMENT NO.7 Dated 31.03.2020

Ref No: HLL/SD/RBD/2019-20/TENDER/15 Dt: 05.03.2020

Title: Essential Technical features for Ventilator for COVID 19

The following amendment has been incorporated to the bid document for the above tender;
The Essential Technical features for Ventilator for COVID 19 is given below. All other terms and conditions as per amendment no. 06 dated 28.03.2020 remains same.

01. The machine should be turbine / compressor based because the installation sites might not have central Oxygen Lines.
02. The machine should have Invasive, Non- Invasive and CPAP features to make them versatile.
03. 200-600 L tidal volume, Lung Mechanics Display
04. Monitoring of plateau Pressure, PEEP, PS, Oxygen Concentration, Lung mechanics /inverse ratio(I:E)
05. Pressure & Volume control & PSV
06. Continuous working capability for 4-5 days

**Senior Manager
Sourcing Division - RBD
HLL Lifecare Ltd.
HLL Bhavan, Poojappura,
Thiruvananthapuram.
Ph.no: 0471 2353932.**