

AMENDMENT NO.1 Dated 16.09.2021

Ref IFB No: HLL/SD/RBD/2021-22/TENDER/08 Dt: 10.09.2021

The following amendment has been incorporated in the above mentioned IFB for Supply of Consumables for onward supplies to Foreign countries (Fiji & Mozambique)

1. Instruction to Bidders (ITB)

FOR

**Section 1, Pont V. Technical Specification,
Annexure 14 – Technical Compliance Sheet**

1. Coverall

- Single use (Disposable), Light colour.
- Impermeable to blood and body fluids, Virus resistant and comfortable to wear.
- Coverall With Breathable fabric Tyvek of 70 GSM resistant to bacterial and viral penetration with integrated boot legging
- The Coverall (including seam) and Shoe Covers for COVID-19 shall conform to the requirements as specified in Table 1 of IS 17423 : 2020 .
- Seams are to be sealed with a Tape of suitable material of medical grade of minimum 16 mm width .
- Coverall size: Universal Fit .
- Integrated hood with elastic around face opening .
- Provided with suitable fastening arrangement which is covered with a storm flap with provision of self-adhesive sealing.
- Zipper of the coverall shall be covered with a flap to avoid accumulation of microbes.
- Soft Elastic to be fitted around Front of hood, wrists & ankles
- Boot Cover:
- Each coverall is provided with a pair of shoe covers with an elastic strip to tighten it, with the Coverall, so that there is no passage for air through it
- Product must be manufactured in India

MAY BE READ US

**Section 1, Pont V. Technical Specification,
Annexure 14 – Technical Compliance Sheet**

1. Coverall

- Single use (Disposable), Light colour.
- Impermeable to blood and body fluids, Virus resistant and comfortable to wear.
- Coverall With **Breathable fabric of 70 GSM** resistant to bacterial and viral penetration with integrated boot legging **with ISI certification**
- The Coverall (including seam) and Shoe Covers for COVID-19 shall conform to the requirements as specified in Table 1 of IS 17423 : 2020 (**latest revision**) .
- **IS 17423 : 2020 certificate to be submitted**
- Seams are to be sealed with a Tape of suitable material of medical grade of minimum 16 mm width .
- Coverall size: Universal Fit .
- Integrated hood with elastic around face opening .
- Provided with suitable fastening arrangement which is covered with a storm flap with provision of self-adhesive sealing.
- Zipper of the coverall shall be covered with a flap to avoid accumulation of microbes.

- Soft Elastic to be fitted around Front of hood, wrists & ankles
- Boot Cover:
- Each coverall is provided with a pair of shoe covers with an elastic strip to tighten it, with the Coverall, so that there is no passage for air through it
- Product must be manufactured in India
- **Recent test report from Government testing lab for Infection barrier test and Breathability test as per IS to be submitted along with the consignment.**

*The above changes will be applicable both in Technical specification and Technical Specification Compliance sheet.

Deputy General Manager (SD-RBD)