

# Response To Pre-Bid Queries (Pre-Bid date: 24.04.2018)

NIB Ref: HITES/PCD/NCI-AIMS/14/18-19 Dated: 13.04.2018

## Schedule No. 07 - Adult and Pediatric Reusable Flexible Intubation Video Endoscope with Screer (3000002783)

Sl. No	Reference to the bidding document	Specification/Clause of the bidding documents	Name of firms	Queries/Clarification asked by firms	Recommendations/replies by the Committee
1	Page no. 66 1	Flexible Intubation Endoscope with CMOS chip on tip for digitally transferring the image to the screen. There should be NO Optical Fiber bundles. Intubation Endoscope to display Full Frame 4:3 Imaging, replacing the circular display windows. The image can be displayed directly on a small TFT monitor attached to the screen	Olympus Medical Systems India Pvt Ltd	Kindly amend it as CCD/CMOS chip on tip & also add compact high resolution Video processor With intergated /seprated LED /Xenon Light source to be quoted from same OEM  Kindly requesting you to delete the line The image can be displayed directly on a monitor Remarks: It is is company specific	No Change
2	Page no. 66 2	Screen 7 inch or more in size for display with clearly arranged soft: keys for immediate use. The screen should have HDMI or HD for connecting to a big screen. Also, there should be port to connect USB drive.	Olympus Medical Systems India Pvt Ltd	Kindly amend it as screen size to be 19-21 inch medica grade with HD inputs and out puts Remarks: It is is company specific	No Change
3			Karl Storz	We Comply.Since you have asked 6No. Of adult scope and 3No. Of paediatric scope therefore we suggest that atleast 3No. Of monitor are required.	Amended as: Screen 7 inch or more for each scope in size for display with clearly arranged soft keys for immediate use. The screen should have HDMI or HD for connecting to a b g screen. Also, there should be port to connect USB drive.
4	Page no. 66 3	Automatic/- manual white balance facility should be available on the monitor as well as on the scope	Olympus Medical Systems India Pvt Ltd	Kindly amend it as Automatic/ manual white balance facility should be available with the system	No Change
5	Page no. 66 4	Documentation of Video & still images should be possible on data card or USB drive with JPEG and MPEG4 format which can be easily transferred to the computer/laptop. Documented videos & still images should be easily recalled on the monitor	Olympus Medical Systems India Pvt Ltd	Kindly delete Remarks: It is is company specific	No Change
6	Page no. 66 5	Screen should be rechargeable & runs on Lithium Ion Batteries (operating time- 1 hr or more)	Olympus Medical Systems India Pvt Ltd	Kindly delete Remarks: It is is company specific	No Change
7	Page no. 66 7	Airway Guide (cum Bite block) for Oral Intubation should be provided with the set (at least 10 airways) .	Olympus Medical Systems India Pvt Ltd	Kindly amend it as mouth piece for oral intubation should be provided Remarks: It is is company specific	No Change
8	Page no. 66 8	Monitor should have a facility to connect flexible scope directly without any special coupler or any additional accessory.	Olympus Medical Systems India Pvt Ltd	Kindly amend it as Video processor should have facility to connect flexible scope directly without any special coupler or any additional accessory. Remarks: It is is company specific	No Change

*[Signature]*

*[Signature]*  
*[Signature]*



Sl. No	Reference to the bidding document	Specification/Clause of the bidding documents	Name of firms	Queries/Clarification asked by firms	Recommendations/Replies by the Committee
9	Page no. 66 9	Integrated recording should be done on a USB drive.	Olympus Medical Systems India Pvt Ltd	Kindly amend it as usb port for still/recording should be present Remarks: It is is company specific	No Change
10	Page no. 66 10	TUBE HOLDER should be a part of standard accessory	Olympus Medical Systems India Pvt Ltd	Kindly amend it as OEM trolley should be provided along with scope hanging pole Remarks: It is is company specific	No Change
11	Page no. 66 11	Set should include- Suction Adaptors (20 Disposable per scope), Cleaning brush ( 20 pieces per scope ) & Leakage tester as standard accessories	Olympus Medical Systems India Pvt Ltd	Kindly amend it as Set should include Disposable Suction valve -20 No's Cleaning brush ( 20 pieces per scope ) & Leakage tester as standard accessories Remarks: It is is company specific	No Change
12	Page no. 66 14 a)	Technical Details of Adult Flexible Video Endoscope- ( 6 Number ) a) Tip deflection UP/DOWN: 140°/140° or more	Olympus Medical Systems India Pvt Ltd	Kindly amend Tip deflection UP/DOWN as 140 /130 degree or more Remarks: It is is company specific	Amended as: Tip deflection UP/DOWN: 140°/130° or more
13	Page no. 66 14 c)	Technical Details of Adult Flexible Video Endoscope- ( 6 Number ) c) Working Length: at least 65 cm,	Olympus Medical Systems India Pvt Ltd	Kindly amend it as atleast 60-65 cm	Amended as: Working Length: at least 60 cm
14	Page no. 66 14 d)	Technical Details of Adult Flexible Video Endoscope- ( 6 Number ) d) Total length: 90 cm or more	Olympus Medical Systems India Pvt Ltd	Kindly amend it as 85 cm or more	Amended as: Total length: 85 cm or more
15	Page no. 66 15 a)	Technical Details of Pediatric Flexible Video Endoscope- ( 3 Number ) a) Tip deflection UP/DOWN: 140°/140° or more	Olympus Medical Systems India Pvt Ltd	Kindly amend Tip deflection UP/DOWN as 140 /130 degree or more	No Change
16	Page no. 66 15 c)	Technical Details of Pediatric Flexible Video Endoscope- ( 3 Number ) Working Length: at least 65 cm,	Olympus Medical Systems India Pvt Ltd	Kindly amend it as atleast 60-65 cm	Amended as: Working Length: at least 60 cm
17	Page no. 66 15 d)	Technical Details of Pediatric Flexible Video Endoscope- ( 3 Number ) Total length: at least 90 cm or more	Olympus Medical Systems India Pvt Ltd	Kindly amend it as 85 cm or more	Amended as: Total length: 85 cm or more
18	Page no. 66 16	In addition, one Pediatric bronchoscope at least 2.5mm must be supplied and should be compatible with existing monitor.	Olympus Medical Systems India Pvt Ltd	Kindly amend it as In addition, one Pediatric bronchoscope at least 2.5mm with channel of 1.2mm & must compatible with video processor & Light source	Amended as: In addition, one Pediatric bronchoscope at least 3.0mm must be supplied and should be compatible with existing monitor
19			Karl Storz	Kindly amend it to neonatal video bronchoscope of size atleast 3mm with working length between 50-55cm compatible with existing monitor.	
20	Page no. 67 19	Computer with software should be supplied with suitable computer system with facility for recording images and video.	Karl Storz	Kindly delete as the recording of image and video in general are with the control of monitor itself.	Deleted

Sl. No.	Reference to the bidding document	Specification/Cause of the bidding documents	Name of firms	Queries/Clarification asked by firms	Recommendations/replies by the Committee
21	Page no. 67 BOQ Sl.no.5	Trolley as per technical specifications - 1 no.	as per prebid discussion	Trolley as per technical specifications - 3 no.	<b>Amended Qty:</b> Trolley as per technical specifications - 3 no.
	Page no. 67 BOQ	Added para	Karl Storz	Kindly add in BOQ the control monitor is 7" with atleast 3No. Of quantity.	<b>BOQ:</b> Added items: Sl.no. 6: Screen 7" or more as per technical specifications - 9 nos. Sl.no.7: Biopsy forceps - 2nos. Sl.no.8: Foreight body removal forceps - 2 nos. Sl.no.9: Brush biopsy forceps - 2nos.

*Ad*

*Geo*

*Swathi*

*Chand*

*Ch*

*Ch*



1

2

3

4

5

6

Name of the person		Date		Time		Place		Remarks	