

Response To Pre-Bid Queries (Pre-Bid date: 24.04.2018)

NIB Ref: HITES/PCD/NCI-AIIMS/14/18-19 Dated: 13.04.2018

Schedule No. 23 (Rfx/Event number 3000002799) Video Laryngoscope for Onco Anesthesia

Sl. No	Reference to the bidding document	Specification/Clause of the bidding documents	Name of firms	Queries/Clarification asked by firms	Recommendations/replies by the Committee
	1	Required is Macintosh blade with closed European Metal finish size 2, 3 and 4 with integrated camera chip and LED light illumination for obtaining more than 50000 Lux of brightness.	Karl Storz	It must be changed to Required is Macintosh blade with closed European Metal finish size 2, 3 and 4 with integrated camera chip and LED light illumination for obtaining good brightness. More than 50000 Lux word need to be removed.	Amended as: Required is Macintosh blade with closed European Metal finish size 2, 3 and 4 with integrated camera chip and LED light illumination

Singh

for

Amend

for