

Amendment No. 3

Date: 05/08/2014

Sub: Amendment to the Tender Enquiry Document

Ref: Items tendered under NIT No. HLL/PCD/GNCTD/08/LNH/14-15 dated 04/06/2014

The following changes have been incorporated in items tendered under the referred NIT.

Section – VII **Technical Specification**

Item Sl. No. 1

Electric Drill Machine (High Speed Electrical Micro Drill)

1. **Existing: Para 8:** The high speed motor should weigh less than 90 grams & should be smaller than 85 mm in length

Amended as: The high speed motor should **weigh less than 250 grams** & should be **smaller than 105 mm in length**

2. **Existing: Para 9:** The motor should have min 110 watts with high torque up to 2.5 Ncm

Amended as: The motor should have min 110 watts with high **torque up to 4.4 Ncm**

3. **Existing: Para 10:** Should have single pedal foot control for varying the speed and forward & reverse rotation

Amended as: Should have **double pedal footswitch in order to have forward, reverse and oscillation mode. Should have provision to control irrigation pump as well.**

4. **Existing: Para 15:** Should have provision to use various saw system

Amended as: Should have provision to use with **oscillating, saggittal and reciprocating saw with at least 14000CPM**

5. **Existing: Para 16:** System should be supplied with reusable perforator cutter for adult use

Amended as: System should be supplied with **disposable perforator, for min 5 surgeries and universal perforator cutter** for adult use.

6. Existing: Para 18: The Craniotomehandpeice should be supplied with fixed dura guard of standard sizes

Amended as: The **disposable** Craniotome handpeice should be supplied with dura guard of **Pediatric, adult and large sizes with cutting routers 5 each for each size.**

All other contents of the tender enquiry including terms & conditions remain unaltered.