

## **TENDER ENQUIRY NOTICE**

HLL/ID/C & M/ KKD-Med Equip-Phase II/2014-15/521

05.06.2015

Name of Work : Supply of Deep Freezer-80 Deg for the Strengthening and Upgradation of Govt: Medical College ,Kozhikode -Phase II.

Limited tenders are invited on behalf of The Principal, Govt: Medical College, Kozhikode for the Supply of Deep Freezer for the Strengthening and Upgradation of Govt: Medical College ,Kozhikode -Phase II. The specification for the items is enclosed herewith.

### **Address to which the offers are to be submitted:**

#### **Deputy Vice President (Technical)**

HLL Lifecare Limited

"ADARSH", TC - 6/1718 (1)

Vettamukku, Thirumala P O

Thiruvananthapuram - 695 006

Ph : 0471 - 2365872/73

Fax: 0471 - 2368144

### **Submission of Offer**

The offer shall be submitted in single bid system. *The offer should be submitted in a sealed envelope super scribing "Supply of Deep Freezer -80 Deg for the Strengthening and Upgradation of Govt: Medical College ,Kozhikode -Phase II.*

**Last date and time of Receipt of offers :15.06.2015 11.00 am.**

**Date and time of Opening of Offers :15.06.2015 11.30 am**

### **Payment terms**

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

**a) On delivery:**

75 % payment of the contract price shall be paid on receipt of goods in good condition and upon the submission of the following documents:

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;

- (ii) Consignee Receipt Certificate as per in original issued by the authorized representative of the consignee;
- (iii) Two copies of packing list identifying contents of each package;
- (iv) Inspection certificate issued by HLL.
- (v) Certificate of origin.

**b) On Acceptance:**

Balance 25 % payment would be made after satisfactory completion against 'Final Acceptance Certificate' of goods/equipments to be issued by the consignees and endorsed by HLL subject to recoveries, if any, either on account of non-rectification of defects/deficiencies not attended by the Supplier or otherwise.

**The address to which the items have to be supplied is as follows:**

The Principal, Govt: Medical College ,Kozhikode  
Kozhikode -673008

**Delivery period**

The Equipments shall be supplied within 1 month from the date of supply order.

**The supplier should give CMC for the quoted item for 5 years after 2 years warranty period.  
Cost of CMC will not be considered for price comparison purpose.**

1	2	3	4	5					6
Sch edu le No.	BRIEF DESCRIPTI ON OF GOODS	QUANT ITY. (Nos.)	Basic price of the equipment excluding tax	CMC Cost for Each Unit year wise*.					Total CMC Contract Cost for 5 Years
				1 <sup>st</sup>	2 <sup>nd</sup>	3 <sup>rd</sup> d	4 <sup>th</sup>	5 <sup>th</sup>	
				a	b	c	d	e	
1									
2									
3									

*The cost of CMC quoted shall be exclusive of service tax . The service tax shall be extra as applicable.*

### **Other Conditions**

1. Warranty should be for minimum 2 years from the date of final acceptance of the equipments and the Warranty has to be given in the name of client.

### **Performance Security**

2. Within Fifteen (15) days from date of the issue of purchase order the successful bidder, shall furnish performance security for an amount equal to ten percent (10%) of the total value of the contract, valid up to sixty (60) days after the date of completion of all contractual obligations by the supplier, including the warranty obligations.
3. Demo of equipment and sample of instruments should be arranged when asked for
4. Earnest Money of Rs 12000/- has to be submitted along with the tender in the form of a Demand Draft/ Fixed Deposit Receipt (FDR) of a scheduled bank issued in favour of HLL Lifecare Limited, Thiruvananthapuram which should be placed in a separate sealed cover marked "Earnest Money" shall be submitted along with the tender.
5. The rate quoted should be inclusive of all taxes, duties, transport, loading and unloading up to consignee site.
6. The tenderer shall submit the compliance for technical specifications. The make and Model should also be attached along with the bid.
7. The original Literature or the relevant part of the user/ service manual should be attached as proof.
8. The offer shall be valid for 3 months.
9. The invoice/ bills shall be raised in the name of "The Principal, Govt: Medical College, Kozhikode a/c HLL Lifecare Ltd" and the same shall be submitted to the HLL Engineer in charge.
10. TIN no. of HLL should not be mentioned in any of the documents as HLL is acting on behalf of the client in the capacity of consultant alone.
11. The completed works shall be handed over to the client after due verification by HLL Engineer in Charge.
12. **The bidders should seal and sign the Tender Enquiry Notice and should be submitted along with the offer.**

**Deputy Vice President (Tech)**

**HLL Lifecare Limited**

<b>5</b>	<b>DEEP FREEZER--80 DEG</b>
<b>5.1</b>	<b>Description of Function</b>
	Deep Freezers are required to preserve blood and blood products, vaccinations etc at specified temperature.
<b>5.2</b>	<b>Operational Requirements</b>
	Internal Minimum Capacity 350 to 400 L net at least double door with at least 6 adjustable shelves.
	Range up to -65 degree Celsius to -85degree celsius(Adjustable)
	Vertical Cabinet (upright model)
<b>5.3</b>	<b>Technical Specifications</b>
	Construction: Solid rust free cabinet to prevent corrosion and lockable castor wheels. Inner surface should be SS304.
	Control System:Micro-processor based temperature controller with digital temperature display LED/LCD with seven days graphic temperature recorder with rechargeable battery back up including charger maintenance free and insensitive to vibration.
	Refrigeration System : Heavy Duty refrigeration system, maintenance free,below -85°C (+ 10 C) with hermetically sealed dual compressors, noise free and vibration free, air cooled with security lock to prevent unintentional switch off shall be supplied.
	Alarm :It should also have audio visual Electronic Alarm System independent of power supply.
	Insulation : High density polyurethane or equivalent Gaskets - Double seal silicon.
	Door heating system for easy opening of door
<b>5.4</b>	<b>System Configuration Accessories, spares and consumables</b>
	As specified
<b>5.5</b>	<b>Environmental factors</b>
	The unit shall be capable of being stored continuously in ambient temperature of 0 - 50deg C and relative humidity of 15-90%
	The unit shall be capable of operating continuously in ambient temperature of 10 - 40deg C and relative humidity of 15-90%
<b>5.6</b>	<b>Power Supply</b>
	Power input to be 220-240VAC, 50Hz fitted with Indian plug
	Suitable Servo controlled Stabilizer/CVT
<b>5.7</b>	<b>Standards and Safety</b>

	Should be FDA or CE or ISI approved product
<b>5.8</b>	<b>Documentation</b>
	User manual in English
	Certificate of calibration and inspection from factory.