

Equipment Specification Data Sheet

Equipment Name: BOD Incubator

Document No.: DS-BIC 02

Revision: 00

Project No.: 120310

Project Name: Integrated Vaccines
Complex, Chengalpattu


Block Code	Block Name	Identification No.	Capacity	Quantity
Q1	Admin QA & QC	Q1-BIC 01	200L to 260L	1


NNE Pharmaplan India Limited


Name	Designation	Signature	Date
Prepared by			
Mr. Sandeep Kumar	Process Engineer	<i>Sandeep</i>	18/02/16
Checked by			
Mr. Yogesha M J	Process Engineer	<i>yogesha-mj</i>	18/02/2016
Approved by			
Dr. Harshad Mali	Lead Engineer - Process	<i>HM</i>	19/02/16

HLL Biotech Limited

Name	Designation	Signature	Date
Reviewed by			
User department: Quality Control <i>SANDEEP KUMAR</i>	Manager	<i>SK</i>	29/02/16
Project / Engineering department <i>A. Anuro</i>	Manager	<i>A. Anuro</i>	29.02.16
Approved by			
Head of the department: Quality Control <i>Sivaganesan</i>	Manager	<i>Plaw</i>	04/03/16
Head of the department (Quality Assurance) <i>A. Suresh Babu</i>	Manager	<i>A. Suresh Babu</i>	04.03.16
Authorized by			
Project Authority <i>Ram K R.</i>	CEO	<i>Lat</i>	7.3.16

Equipment Specification Data Sheet		
HLL Biotech Limited, CHENNAI		
nne pharmaplan®	INTEGRATED VACCINES COMPLEX, CHENGALPATTU	
	Equipment Name	BOD Incubator
	Document No.	DS-BIC 02
	Revision No.	00
		
1	Process Requirement	
1.1	The incubator shall be used for storage of samples under controlled conditions.	
2	Equipment ID	
2.1	Q1-BIC 01	
3	Technical Specification	
3.1	Model	BOD Incubator
3.2	Temperature control	Microprocessor with P.I.D
3.3	Temperature range	15 to 60 °C
3.4	Temperature Readability(Resolution)	0.1 °C
3.5	Number of shelves	3 or more adjustable and perforated SS304 shelves
3.6	Refrigerant	CFC Free
3.7	Display type	LED/LCD Display
3.8	Capacity	200L to 260L
3.9	Power supply	To be compatible to standard Indian Power supply socket
3.10	Quantity	1 No
3.11	Dimensions (W x D x H) Internal Work area External dimensions	Vendor to specify
4	Material of Construction	
4.1	External body Construction	cGMP compalint epoxy coated material
4.2	Internal body Construction	SS 316L/SS 304, electropolished.
4.3	Inner Door	Safety transparent door
4.4	Outer Door	SS304/Powder coated
4.5	Finishes	a) Rounded inner chamber corners for easy cleaning b) Design of the equipment should enhance cleaning by providing minimum sharp corners, minimum crevices and smooth finished welds joints
4.6	Gaskets, seals, O-ring	Food Grade/ nontoxic material like neoprene or better. Use of Asbestos is prohibited
4.7	Expected operational hours per day	24 hrs
4.8	Validation	Validation port to be provided to insert probes for temperature mapping
5	Specific Equipment Requirements	
5.1	Temperature mapping to be provided at the time of installation	
5.2	The body construction shall provide tempearture stability.	
5.3	The gap between inner & outer walls of chamber should be fitted with high grade polyurathane foam (PUF), to ensure maximum temperature efficiency.	
5.4	It shall have forced air circulation to achieve uniformity of conditions.	
5.5	Incubator shall include a safety thermostat.	
5.6	Compressor shall have overload relay.	
5.7	There should be independent over-temperature safety protection.	
5.8	It should have lockable adjustable feet and stable casters for stability and easy repositioning incubator.	
5.9	Alarm : (Visual - Audio) for temperature deviation	

Equipment Specification Data Sheet		
HLL Biotech Limited, CHENNAI		
nne pharmaplan*	INTEGRATED VACCINES COMPLEX, CHENGALPATTU	
	Equipment Name	BOD Incubator
	Document No.	DS-BIC 02
	Revision No.	00
		
5.10	Equipment design must realize zero contamination.	
5.11	The heat/noise level given off by the unit shall be stated (inside the room).	
5.12	The equipment shall be compatible for cleaning with all standard disinfectants	
6	Other Requirement	
6.1	Training/Demo for the users on operation and cleaning to be provided.	
7	Regulatory aspects	
7.1	CE certification	
8	Safety Requirements	
	Following facilities must be provided to protect personnel and equipment:	
8.1	Appropriate closure of all parts	
8.2	Proper earthing is necessary	
8.3	Doors interlocking alarm (visual/ audio).	
8.4	Noise level should be below 60 decible at a distance of 1m from the equipment	
8.5	Chamber shall be insulated properly to maintain inner environment	
9	Documents	
9.1	Following documents, but not limited to these, are expected from the vendor as part of the supply package in hard copy as well as editable electronic file	
9.2	IOQ document.	
9.3	Operation and maintenance manuals shall be provided along with IOQ documents during installation at site	
9.4	Calibration certificate of critical instruments with respect to the traceable national reference standard instrument and their calibration procedure.	
9.5	Warranty Letter for Minimum 1 year from the date of installation.	
9.6	Vendor should provide list of standard spare parts with ordering information.	
9.7	Vendor should provide list of change parts (if applicable) with ordering information	
10	Timelines	
10.1	Not Applicable	
11	Preferred list of Makes	
11.1	Binder, Memmert, Thermo Scientific, Eppendorf	
	NOTE: Accurate size and technical specification need to be mentioned by the vendor	

Equipment Specification Data Sheet					
HLL Biotech Limited, CHENNAI					
nne pharma plan*	INTEGRATED VACCINES COMPLEX, CHENGALPATTU			 HLL BIOTECH LIMITED <small>(A Subsidiary of HLL Life Sciences Limited) (A Government of India Enterprise)</small>	
	Equipment Name	BOD Incubator			
	Document No.	DS-BIC 02			
	Revision No.	00			
Table-1: Equipment location					
Equipment ID	Block Name	Room Name	Room No	Room dimension in mm	Room height in mm
Q1-BIC 01	Admin QA & QC	Microbiology Lab	Q1S045	3700 x 4600	2400
Table-2: Change Log					
Date	Name	Revision	Section	Change/Comment	
10-02-2016	Sandeep Kumar	00	-	New document	
Table-3: Annexure					
Not applicable					