

## **TENDER ENQUIRY NOTICE**

HLL/ID/Alpza-Med Equip/2014-15 /

21.11.2014

Name of Work : Supply of 1 no: of Non Invasive Cardiac output Monitor for the setting up of emergency and critical care department at Alappuzha Medical College.

Limited tenders are invited on behalf of Principal, Alappuzha Medical College, Govt. of Kerala for the Supply of 1 no: of Non Invasive Cardiac output Monitor for the setting up of emergency and critical care department at Alappuzha Medical College. The specification for the Non Invasive Cardiac output Monitor is enclosed herewith.

### **Address to which the offers are to be submitted:**

#### **Deputy Vice President (Technical)**

HLL Lifecare Limited

“ADARSH”, TC - 6/1718 (1)

Vettamukku, Thirumala P O

Thiruvananthapuram - 695 006

Ph : 0471 - 2365872/73

Fax: 0471 - 2368144

### **Submission of Offer**

The offer shall be submitted in single bid system. The offer should be submitted in a sealed envelope superscribing “Supply of 1 no: of Non Invasive Cardiac output Monitor for the setting up of emergency and critical care department at Alappuzha Medical College.”

Last date and time of Receipt of offers : 29.11.2014 11.00 am.

Date and time of Opening of Offers : 29.11.2014 11.30 am

### **Payment terms**

Payment shall be made in Indian Rupees as specified in the contract in the following manner:

**a) On delivery:**

75 % payment of the contract price shall be paid on receipt of goods in good condition and upon the submission of the following documents:

- (i) Four copies of supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
- (ii) Consignee Receipt Certificate as per in original issued by the authorized representative of the consignee;
- (iii) Two copies of packing list identifying contents of each package;
- (iv) Inspection certificate issued by HLL.
- (v) Certificate of origin.

**b) On Acceptance:**

Balance 25 % payment would be made after satisfactory completion against 'Final Acceptance Certificate' of goods/equipments to be issued by the consignees and endorsed by HLL subject to recoveries, if any, either on account of non-rectification of defects/deficiencies not attended by the Supplier or otherwise.

### **The address to which the Non Invasive Cardiac output Monitor have to be supplied is as follows:**

The Principal, Govt Medical College Alappuzha  
Vandanam  
Alappuzha- 688005

### **Delivery period**

The **Non Invasive Cardiac output Monitor** shall be supplied within 1 month from the date of supply order.

The supplier should give CMC for the quoted item for 5 years after 2 years warranty period. Cost of CMC will not be considered for price comparison purpose.

1	2	3	4					5
Schedule No.	BRIEF DESCRIPTION OF GOODS	QUANTITY. (Nos.)	CMC Cost for Each Unit year wise*.					Total CMC Contract Cost for 5 Years
			1 <sup>st</sup>	2 <sup>nd</sup>	3 <sup>rd</sup>	4 <sup>th</sup>	5 <sup>th</sup>	
			a	b	c	d	e	
1								
2								
3								

### **Other Conditions**

1. Warranty should be for minimum 2 years from the date of acceptance of Non Invasive Cardiac output Monitor and the Warranty has to be given in the name of client.

#### **Performance Security**

2. Within Fifteen (15) days from date of the issue of purchase order the successful bidder, shall furnish performance security for an amount equal to ten percent (10%) of the total value of the contract, valid up to sixty (60) days after the date of completion of all contractual obligations by the supplier, including the warranty obligations.

3. Demo of equipments and sample of instruments should be arranged when asked for

4. Earnest Money of Rs. 14,000/- has to be submitted along with the tender in the form of a Demand Draft/ Fixed Deposit Receipt (FDR) of a scheduled bank issued in favour of HLL Lifecare Limited, Thiruvananthapuram which should be placed in a separate sealed cover marked "Earnest Money" shall be submitted along with the tender.

5. The rate quoted should be inclusive of all taxes, duties, transport, loading and unloading up to consignee site.

6. The tenderer shall submit the compliance for technical specifications. The make and Model should also be attached along with the bid.

7. The original Literature or the relevant part of the user/service manual should be attached as proof.
8. The offer shall be valid for 3 months.
9. The invoice/ bills shall be raised in the name of “**Principal, Alappuzha Medical College a/c HLL Lifecare Ltd**” and the same shall be submitted to the HLL Engineer in charge.
10. TIN no. of HLL should not be mentioned in any of the documents as HLL is acting on behalf of the client in the capacity of consultant alone.
11. The completed works shall be handed over to the client after due verification by HLL Engineer in Charge.

**Deputy Vice President (Tech)**  
**HLL Lifecare Limited**

	<b>MULTIPARAMETER MONITOR WITH NON INVASIVE CARDIAC OUTPUT</b>
<b>1</b>	<b>Technical Specifications</b>
1.1	Parameters required - ECG, Respiration, NIBP, SpO2, Temperature, and <b>Non Invasive Cardiac Output</b>
1.2	Should have eight digital and waveforms/traces display
1.3	Should have minimum 12 inches multicoloured LCD display screen.
1.4	Should have the capability for storage of patient data and printing of patient reports.
1.5	Monitor should have audible and visual alarms capability. .
1.6	Should have automatic arrhythmia detection & alarm for standard and lethal arrhythmia.
1.7	Should have trend of at least 24 hours.
1.8	The monitors should have battery back up of minimum 2 hrs
1.9	Power input to be 220-240VAC, 50Hz fitted with Indian plug
1.10	Monitors should be supplied with wallmount, swivel type stands of high quality
<b>2</b>	<b>System Configuration Accessories, spares and consumables for Bedside monitor</b>
2.1	ECG/Resp :5 Lead ECG Cable with clip- 1 sets per monitor
2.2	NIBP:Adult cuff- with cable -01 nos. per monitor
2.3	SpO2:Adult SpO2 sensor with cable- one nos per monitor
2.4	Temperature: Rectal temperature probe- one per monitor and skin temperature probe- one per monitor
2.5	Should be supplied with all accessories and consumables required for measurement of <b>non invasive cardiac output</b> .
<b>2</b>	<b>Standards, Safety and Training</b>
2.1	Manufacturer should have ISO certification for quality standards and copy of the certificate should be submitted along with the technical bid.
2.2	Product should be US FDA/ European CE approved
2.2	Comprehensive training should be given for staffs and engineers till familiar with the system.
<b>3</b>	<b>Documentation</b>
3.1	Must submit atleast 2 nos of latest purchase order of the quoted model dated within 1 year along with the <b>price bid</b> .